

Bovine Von Willebrand Factor ELISA Kit | VWF elisa kit

Bovine Von Willebrand Factor ELISA Kit

CATALOG #
AAA10041

REACTIVITY
Bovine


PRODUCT SYNONYMS

Von Willebrand Factor; Bovine Von Willebrand Factor ELISA Kit; VWF elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Bovine
Specificity	No significant cross-reactivity or interference between Bovine VWF and analogues was observed.
Assay Type	Quantitative Sandwich
Detection Range	62.5U/L-2000U/L
Sensitivity	10U/L
Intra-assay Precision	Intra-assay CV (%) is less than 15%.
Inter-assay Precision	Inter-assay CV (%) is less than 15%. [CV(%) = SD/mean ×100]. All CV% should be compared by concentration, not compared by OD values.
Preparation and Storage	Store all reagents at 2-8 degree C

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for VWF elisa kit

Background: This Quantitative Sandwich ELISA kit is for lab reagent/research use only, not for drug, household, therapeutic or diagnostic applications! This kit is intended to be used for determine the level of VWF (hereafter termed "analyte") in undiluted original Bovine serum, plasma or tissue homogenates samples. For other sample types please contact tech support to determine compatibility with this assay. This kit is not suitable for assaying non-biological sources of substances.

NCBI AND UNIPROT INFORMATION

NCBI GI #	89191868
NCBI GeneID	7450

NCBI Accession #	NP_000543.2
NCBI GenBank Nucleotide #	NM_000552.3
UniProt Accession #	P04275; Q99806
Molecular Weight	309,265 Da
NCBI Official Full Name	von Willebrand factor preproprotein
NCBI Official Synonym Full Names	von Willebrand factor
NCBI Official Symbol	VWF
NCBI Official Synonym Symbols	VWD; F8VWF
NCBI Protein Information	von Willebrand factor; coagulation factor VIII VWF
UniProt Protein Name	von Willebrand factor
UniProt Gene Name	VWF
UniProt Synonym Gene Names	F8VWF; vWF
UniProt Entry Name	VWF_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/vwf-bovine-ovine-elisa-kits/von-willebrand-factor/10041>