

Human Beta-site APP Cleaving Enzyme 1 ELISA Kit | BACE1 elisa kit

Human Beta-site APP Cleaving Enzyme 1 ELISA Kit

CATALOG #
AAA100936

REACTIVITY
Human


PRODUCT SYNONYMS

Beta-site APP Cleaving Enzyme 1; Human Beta-site APP Cleaving Enzyme 1 ELISA Kit; BACE1/APP beta-secretase/ASP2/Aspartyl protease 2/BACEAsp 2/beta-secretase 1/beta-secretase 1 precursor variant 1/beta-site amyloid beta A4 precursor protein-cleaving enzyme/Beta-site amyloid precursor protein cleaving enzyme 1/Beta-site APP cleaving enzyme 1/beta-site APP-cleaving enzyme/beta-site APP-cleaving enzyme 1/EC 3.4.23/EC 3.4.23.46/FLJ90568/HSPC104/KIAA1149/memapsin-2/Memapsin-2/Membrane-associated aspartic protease 2/transmembrane aspartic proteinase Asp2; BACE1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Assay Type	Sandwich ELISA, Double Antibody
Samples	cell or tissue lysate, Other liquid samples
Detection Range	0.156-10ng/ml
Sensitivity	0.094ng/ml
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Standard Curve (Sample)

Typical Data & Standard Curve

Results of a typical standard run of a Human BACE1 ELISA Kit are shown below. This standard curve was generated at our lab for demonstration purpose only. Each user should obtain their own standard curve as per experiment. (N/A=not applicable)

X	ng/ml	0	0.625	1.25	2.5	5	10	20	40
Y	OD450	0.028	0.064	0.095	0.162	0.296	0.687	1.198	2.349

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for BACE1 elisa kit

Principle of the Assay||This kit was based on sandwich enzyme-linked immune-sorbent assay technology. anti-Human BACE1 antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-Human BACE1 antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and wash with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the Human BACE1 amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of Human BACE1 can be calculated.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/bace1-human-elisa-kits/beta-site-app-cleaving-enzyme-1/100936>