

Rat Coagulation Factor XI ELISA Kit | FXI elisa kit

Rat Coagulation Factor XI ELISA Kit

CATALOG #
AAA101101

REACTIVITY
Rat

PRODUCT SYNONYMS

Coagulation Factor XI; Rat Coagulation Factor XI ELISA Kit; FXI; FXI elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This assay has high sensitivity and excellent specificity for detection of FXI . No significant cross-reactivity or interference between FXI and analogues was observed.
Sequence Length	556
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	0.156-10ng/ml
Sensitivity	0.094ng/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level FXI were tested 20 times on one plate, respectively. Intra-Assay: CV<8%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level FXI were tested on 3 different plates, 8 replicates in each plate. CV (%) = SD/meanX100. Inter-Assay: CV<10%
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES

Typical Data & Standard Curve

Results of a typical standard run of a FXI ELISA Kit are shown below. This standard curve was generated at our lab for demonstration purpose only. Each user should obtain their own standard curve as per experiment. (N/A-not applicable)

X	ng/ml	0	0.156	0.312	0.625	1.25	2.5	5	10
Y	OD450	0.034	0.113	0.183	0.287	0.592	0.973	1.543	2.501

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for FXI elisa kit

Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti- FXI antibody was pre-coated onto 96-well plates. And the biotin conjugated anti- FXI antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and wash with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the FXI amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of FXI can be calculated.

NCBI AND UNIPROT INFORMATION

NCBI GI #	114145441
NCBI GeneID	290757
NCBI Accession #	NP_001041313.1
NCBI GenBank Nucleotide #	NM_001047848.1
NCBI Official Full Name	coagulation factor XI
NCBI Official Synonym Full Names	coagulation factor XI
NCBI Official Symbol	F11
NCBI Official Synonym Symbols	LRRGT00086
NCBI Protein Information	coagulation factor XI

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/fxi-rat-elisa-kits/coagulation-factor-xi/101101>