

Mouse Diamine Oxidase ELISA Kit | DAO elisa kit

Mouse Diamine Oxidase ELISA Kit

CATALOG #
AAA101249

REACTIVITY
Mouse


PRODUCT SYNONYMS

Diamine Oxidase; Mouse Diamine Oxidase ELISA Kit; DAO/AOC1/DAO/Diamine Oxidase/Histaminase/KAO/ABP/amiloride binding protein 1 (amine oxidase (copper-containing))/Amiloride-binding protein/amiloride-sensitive amine oxidase/amiloride-sensitive amine oxidase [copper-containing]/AOC1DAOamiloride-binding protein-1/DAO1/Diamine oxidase/EC 1.4.3/EC 1.4.3.22/Histaminase/histaminase/KAO/Kidney amine oxidase; DAO elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Mouse
Sequence Length	677
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	1.563-100ng/ml
Sensitivity	0.938ng/ml
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Standard Curve (Sample)

Typical Data & Standard Curve

Results of a typical standard run of a Mouse DAO ELISA Kit are shown below. This standard curve was generated at our lab for demonstration purpose only. Each user should obtain their own standard curve as per experiment. (N/A=not applicable)

X	ng/ml	0	3.9	7.8	15.6	31.3	62.5	125	250
Y	OD450	0.017	0.121	0.261	0.485	0.745	1.437	2.091	2.733

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for DAO elisa kit

Principle of the Assay||This kit was based on sandwich enzyme-linked immune-sorbent assay technology. anti-Mouse DAO antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-Mouse DAO antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and wash with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the Mouse DAO amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of Mouse DAO can be calculated.

NCBI AND UNIPROT INFORMATION

NCBI GI #	983174779
NCBI Official Full Name	diamine oxidase

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/dao-mouse-elisa-kits/diamine-oxidase/101249>