

Human Receptor II for the Fc Fragment of IgG ELISA Kit | CD32 elisa kit

Human Receptor II for the Fc Fragment of IgG ELISA Kit

CATALOG # AAA101891	REACTIVITY Human
-------------------------------	----------------------------


PRODUCT SYNONYMS

Receptor II for the Fc Fragment of IgG; Human Receptor II for the Fc Fragment of IgG ELISA Kit; CD32/FcgammaR II (Receptor II for the Fc Fragment of IgG); CD32 elisa kit


PRODUCT SPECIFICATIONS

Reactivity	Human
Sequence Length	293
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	46.875-3000pg/ml
Sensitivity	28.125pg/ml
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Standard Curve (Sample)



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CD32 elisa kit	Principle of the Assay This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti-Human CD32 antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-Human CD32 antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and wash with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the Human CD32 amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of Human CD32can be calculated.
--	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	117616286
NCBI Official Full Name	CD32

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

