

Rat Diamine Oxidase ELISA Kit | DAO elisa kit

Rat Diamine Oxidase ELISA Kit

CATALOG #
AAA102005

REACTIVITY
Rat


PRODUCT SYNONYMS

Diamine Oxidase; Rat Diamine Oxidase ELISA Kit; DAO; DAO elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This assay has high sensitivity and excellent specificity for detection of DAO . No significant cross-reactivity or interference between DAO and analogues was observed.
Sequence Length	677
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	3.125-200ng/ml
Sensitivity	1.875ng/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level DAO were tested 20 times on one plate, respectively. Intra-Assay: CV<8%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level DAO were tested on 3 different plates, 8 replicates in each plate. CV (%) = SD/meanX100. Inter-Assay: CV<10%
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for DAO elisa kit

Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti-DAO antibody was pre-coated onto 96-well plates. And the biotin conjugated anti- DAO antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and washed with wash buffer. HRPStreptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the DAO amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of DAO can be calculated.

NCBI AND UNIPROT INFORMATION

NCBI GI #	983174779
NCBI Official Full Name	diamine oxidase

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/dao-rat-elisa-kits/diamine-oxidase/102005>