

Human Acetylated p53 (Acetylated p53) ELISA Kit

Human Acetylated p53 (Acetylated p53)

CATALOG #
AAA102622

REACTIVITY
Human


PRODUCT SYNONYMS

Acetylated p53 (Acetylated p53); Human Acetylated p53 (Acetylated p53); Acetylated p53 (Acetylated p53) elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Acetylated p53 . No significant cross-reactivity or interference between Acetylated p53 and analogues was observed. Note: Limited by current skills and knowledge, it is difficult for us to complete the crossreactivity detection between Acetylated p53 and all the analogues, therefore, cross reaction may still exist.
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	46.875-3000pg/ml
Sensitivity	28.125pg/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Acetylated p53 were tested 20 times on one plate, respectively. Intra-Assay: CV<8%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Acetylated p53 were tested on 3 different plates, 8 replicates in each plate. CV (%) = SD/meanX100. Inter-Assay: CV<10%
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for Acetylated p53 (Acetylated p53) elisa kit

Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti-Acetylated p53 antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-Acetylated p53 antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and washed with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the Acetylated p53 amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of Acetylated p53 can be calculated.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/acetylated-p53-acetylated-p53/102622>