

Rabbit KGF/FGF7 (Keratinocyte Growth Factor) ELISA Kit | KGF/FGF7 elisa kit

Rabbit KGF/FGF7 (Keratinocyte Growth Factor)

CATALOG #
AAA102658

REACTIVITY
Rabbit

PRODUCT SYNONYMS

KGF/FGF7 (Keratinocyte Growth Factor); Rabbit KGF/FGF7 (Keratinocyte Growth Factor); FGF-7/HBGF-7; KGF/FGF7 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rabbit
Specificity	This assay has high sensitivity and excellent specificity for detection of KGF/FGF7 . No significant cross-reactivity or interference between KGF/FGF7 and analogues was observed.
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	31.25-2000pg/ml
Sensitivity	18.75pg/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level KGF/FGF7 were tested 20 times on one plate, respectively. Intra-Assay: CV<8%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level KGF/FGF7 were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<10%
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES

Typical Data & Standard Curve

Results of a typical standard run of a KGF/FGF7 ELISA Kit are shown below. This standard curve was generated at our lab for demonstration purpose only. Each user should obtain their own standard curve as per experiment. (N/A-not applicable)

X	pg/ml	0	31.25	62.5	125	250	500	1000	2000
Y	OD450	0.051	0.123	0.179	0.276	0.525	0.705	1.337	1.934

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for KGF/FGF7 elisa kit

Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti-KGF/FGF7 antibody was pre-coated onto 96-well plates. And the biotin conjugated anti- KGF/FGF7 antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and wash with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the KGF/FGF7 amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of KGF/FGF7 can be calculated.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.