

Bovine Complement fragment 5a, C5a ELISA Kit | C5a elisa kit

Bovine Complement fragment 5a, C5a ELISA Kit

CATALOG #
AAA11328

REACTIVITY
Bovine


PRODUCT SYNONYMS

Complement fragment 5a; C5a; Bovine Complement fragment 5a; C5a ELISA Kit; C5a elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Bovine
Assay Type	Quantitative Sandwich
Samples	Serum, plasma, cell culture supernates, Ascites, tissue homogenates or other biological fluids
Detection Range	2-600ug/ml
Sensitivity	1.14ug/ml
Intra-assay Precision	Intra-Assay Precision (Precision within an assay) Three samples of known concentration were tested on one plate to assess intra-assay precision. Intra-Assay: CV<8%
Inter-assay Precision	Inter-Assay Precision (Precision between assays) Three samples of known concentration were tested in separate assays to assess inter-assay precision. CV(%) = SD/mean x 100. Inter-Assay: CV<10%
Preparation and Storage	Store at 2 to 8 degree C for 6 months.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for C5a elisa kit

Principle of the Assay: This kit is an Enzyme-Linked Immunosorbent Assay (ELISA). The plate has been pre-coated with Bovine C5a antibody. C5a present in the sample is added and binds to antibodies coated on the wells. And then biotinylated Bovine C5a Antibody is added and binds to C5a in the sample. Then Streptavidin-HRP is added and binds to the Biotinylated C5a antibody. After incubation unbound Streptavidin-HRP is washed away during a washing step. Substrate solution is then added and color develops in proportion to the amount of Bovine C5a. The reaction is terminated by addition of acidic stop solution and absorbance is measured at 450 nm.

Intended Uses: This sandwich kit is for the accurate quantitative detection of Bovine Complement fragment 5a (also known as C5a) in serum, plasma, cell culture supernates, Ascites, tissue homogenates or other biological fluids.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/bovine-ovine-elisa-kits/complement-fragment-5a-c5a/11328>