

Sheep Tri-iodothyronine (T3) ELISA Kit | T3 elisa kit

Sheep Tri-iodothyronine (T3) ELISA Kit

CATALOG #
AAA122087

REACTIVITY
Sheep


PRODUCT SYNONYMS

Tri-iodothyronine (T3); Sheep Tri-iodothyronine (T3) ELISA Kit; T3 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Sheep
Samples	Serum, plasma, cell culture supernates, ascites, tissue homogenates or other biological fluids
Assay Type	Quantitative Sandwich
Detection Range	0.2-60nmol/L
Sensitivity	0.113nmol/L
Intra-assay Precision	Intra-Assay Precision (Precision within an assay) Three samples of known concentration were tested on one plate to assess intra-assay precision. Intra-Assay: CV<8%
Inter-assay Precision	Inter-Assay Precision (Precision between assays) Three samples of known concentration were tested in separate assays to assess inter-assay precision. CV(%) = SD/mean x 100. Inter-Assay: CV<10%
Preparation and Storage	Store at 2 to 8 degree C for 6 months.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for T3 elisa kit

Intended Uses: This sandwich kit is for the accurate quantitative detection of Sheep Tri-iodothyronine (also known as T3) in serum, plasma, cell culture supernates, Ascites, tissue homogenates or other biological fluids.

Principle of the Assay: This kit is an Enzyme-Linked Immunosorbent Assay (ELISA). The plate has been pre-coated with Sheep T3 antibody. T3 present in the sample is added and binds to antibodies coated on the wells. And then biotinylated Sheep T3 Antibody is added and binds to T3 in the sample. Then Streptavidin-HRP is added and binds to the Biotinylated T3 antibody. After incubation unbound Streptavidin-HRP is washed away during a washing step. Substrate solution is then added and color develops in proportion to the amount of Sheep T3. The reaction is terminated by addition of acidic stop solution and absorbance is measured at 450 nm.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

