

Horse Melatonin (MT) ELISA Kit | MT elisa kit

Horse Melatonin (MT) ELISA Kit

CATALOG #
AAA123035

REACTIVITY
Horse


PRODUCT SYNONYMS

Melatonin (MT); Horse Melatonin (MT) ELISA Kit; MT elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Horse
Samples	Serum, plasma, cell culture supernates, ascites, tissue homogenates or other biological fluids
Assay Type	Quantitative Sandwich
Detection Range	7-1500ng/L
Sensitivity	3.74ng/L
Intra-assay Precision	Intra-Assay Precision (Precision within an assay) Three samples of known concentration were tested on one plate to assess intra-assay precision. Intra-Assay: CV<8%
Inter-assay Precision	Inter-Assay Precision (Precision between assays) Three samples of known concentration were tested in separate assays to assess inter-assay precision. CV(%) = SD/mean x 100. Inter-Assay: CV<10%
Preparation and Storage	Store at 2 to 8 degree C for 6 months.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for MT elisa kit

Intended Uses: This sandwich kit is for the accurate quantitative detection of Horse Melatonin (also known as MT) in serum, plasma, cell culture supernates, Ascites, tissue homogenates or other biological fluids.

Principle of the Assay: This kit is an Enzyme-Linked Immunosorbent Assay (ELISA). The plate has been pre-coated with Horse MT antibody. MT present in the sample is added and binds to antibodies coated on the wells. And then biotinylated Horse MT Antibody is added and binds to MT in the sample. Then Streptavidin-HRP is added and binds to the Biotinylated MT antibody. After incubation unbound Streptavidin-HRP is washed away during a washing step. Substrate solution is then added and color develops in proportion to the amount of Horse MT. The reaction is terminated by addition of acidic stop solution and absorbance is measured at 450 nm.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/horse-elisa-kits/melatonin-mt/123035>