

IK Cytokine, Down Regulator Of HLA II (IK) CLIA Kit | IK clia kit

CLIA Kit for IK Cytokine, Down Regulator Of HLA II (IK)

CATALOG #
AAA129941


PRODUCT SYNONYMS

IK Cytokine; Down Regulator Of HLA II (IK); CLIA Kit for IK Cytokine; IK clia kit

PRODUCT SPECIFICATIONS

Specificity	This assay has high sensitivity and excellent specificity for detection of IK Cytokine, Down Regulator Of HLA II (IK). No significant cross-reactivity or interference between IK Cytokine, Down Regulator Of HLA II (IK) and analogues was observed.
Sequence Length	557
Assay Type	Sandwich
Samples	Tissue homogenates and other biological fluids
Detection Range	0.027-20ng/mL. The standard curve concentrations used for the ELISA's were 20ng/mL, 6.67ng/mL, 2.22ng/mL, 0.74ng/mL, 0.247ng/mL, 0.082ng/mL, 0.027ng/mL.
Sensitivity	The minimum detectable dose of this kit is typically less than 0.115ng/mL.
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level IK Cytokine, Down Regulator Of HLA II (IK) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level IK Cytokine, Down Regulator Of HLA II (IK) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = SD/mean \times 100$ Intra-Assay: $CV < 10\%$ Inter-Assay: $CV < 12\%$
Organism Species	Homo sapiens (Human)
Assay Length	4.5 hours
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for IK clia kit	The microtiter plate provided in this kit has been pre-coated with an antigen. Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated secondary antibody. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the IK Cytokine, Down Regulator Of HLA II (IK) level in the sample or standard.
---	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	47940619
NCBI GeneID	3550
UniProt Accession #	Q13123; Q6IPD8
Molecular Weight	65,602 Da
NCBI Official Full Name	IK cytokine, down-regulator of HLA II
NCBI Official Synonym Full Names	IK cytokine, down-regulator of HLA II
NCBI Official Symbol	IK
NCBI Official Synonym Symbols	RED; CSA2
NCBI Protein Information	protein Red; IK factor; RD element; cytokine IK; prer protein; chondrosarcoma-associated protein 2
UniProt Protein Name	Protein Red
UniProt Gene Name	IK
UniProt Synonym Gene Names	RED; RER
UniProt Entry Name	RED_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/clia-kits/ik-cytokine-down-regulator-of-hla-ii-ik/129941>