

Rabbit Serum Amyloid A (SAA) CLIA Kit | SAA clia kit

Rabbit Serum Amyloid A (SAA) CLIA Kit

CATALOG #
AAA130065

REACTIVITY
Rabbit


PRODUCT SYNONYMS

Serum Amyloid A (SAA); Rabbit Serum Amyloid A (SAA) CLIA Kit; SA-A; SAA1; PIG4; TP53I4; Amyloid protein A; Amyloid fibril protein AA; Serum amyloid A-1 protein; SAA clia kit

PRODUCT SPECIFICATIONS

Reactivity	Rabbit
Specificity	This assay has high sensitivity and excellent specificity for detection of Serum Amyloid A (SAA). No significant cross-reactivity or interference between Serum Amyloid A (SAA) and analogues was observed.
Sequence Length	129
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma and other biological fluids
Detection Range	2.74-2000ng/mL
Sensitivity	1.02ng/mL
Application	Chemiluminescent immunoassay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Serum Amyloid A (SAA) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Serum Amyloid A (SAA) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for SAA clia kit	The microplate provided in this kit has been pre-coated with an antibody specific to Serum Amyloid A (SAA). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Serum Amyloid A (SAA). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Serum Amyloid A (SAA) level in the sample or standard.;
Product Categories/Family for SAA clia kit	Infection immunity; Kidney biomarker; Hepatology

NCBI AND UNIPROT INFORMATION

NCBI GI #	93352563
NCBI GeneID	678660
NCBI Accession #	NP_001035288.1
NCBI GenBank Nucleotide #	NM_001040198.1
NCBI Official Full Name	serum amyloid A protein
NCBI Official Symbol	SAA1
NCBI Official Synonym Symbols	SAA
NCBI Protein Information	serum amyloid A protein

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/rabbit-clia-kits/serum-amyloid-a-saa/130065>