

Human Apolipoprotein A4 (APOA4) CLIA Kit | APOA4 clia kit

Human Apolipoprotein A4 (APOA4) CLIA Kit

CATALOG #
AAA130155

REACTIVITY
Human


PRODUCT SYNONYMS

Apolipoprotein A4 (APOA4); Human Apolipoprotein A4 (APOA4) CLIA Kit; ApoA-IV; ApoAIV; Apo-A4; APOA4 clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Apolipoprotein A4 (APOA4). No significant cross-reactivity or interference between Apolipoprotein A4 (APOA4) and analogues was observed.
Sequence Length	396
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma and other biological fluids
Detection Range	0.21-150ng/mL
Sensitivity	0.08ng/mL
Application	Chemiluminescent immunoassay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Apolipoprotein A4 (APOA4) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Apolipoprotein A4 (APOA4) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for APOA4 clia kit	The microplate provided in this kit has been pre-coated with an antibody specific to Apolipoprotein A4 (APOA4). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Apolipoprotein A4 (APOA4). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Apolipoprotein A4 (APOA4) level in the sample or standard.;
Product Categories/Family for APOA4 clia kit	Metabolic pathway; Endocrinology; Cardiovascular biology; Hepatology; Gastroenterology; Nutrition metabolism

NCBI AND UNIPROT INFORMATION

NCBI GI #	71773110
NCBI GeneID	337
NCBI Accession #	NP_000473.2
NCBI GenBank Nucleotide #	NM_000482.3
UniProt Accession #	P06727; Q14CW8; Q6Q787; A8MSL6
Molecular Weight	45,399 Da
NCBI Official Full Name	apolipoprotein A-IV
NCBI Official Synonym Full Names	apolipoprotein A-IV
NCBI Official Symbol	APOA4
NCBI Protein Information	apolipoprotein A-IV
UniProt Protein Name	Apolipoprotein A-IV
UniProt Gene Name	APOA4
UniProt Synonym Gene Names	Apo-AIV; ApoA-IV
UniProt Entry Name	APOA4_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-clia-kits/apolipoprotein-a4-apoa4/130155>