

Human Ubiquitin Conjugating Enzyme E2L3 (UBE2L3) CLIA Kit | UBE2L3 clia kit

Human Ubiquitin Conjugating Enzyme E2L3 (UBE2L3) CLIA Kit

CATALOG #
AAA130367

REACTIVITY
Human


PRODUCT SYNONYMS

Ubiquitin Conjugating Enzyme E2L3 (UBE2L3); Human Ubiquitin Conjugating Enzyme E2L3 (UBE2L3) CLIA Kit; L-UBC; E2-F1; UBCH7; UbcM4; Ubiquitin carrier protein L3; Ubiquitin-conjugating enzyme E2-F1; Ubiquitin-protein ligase L3; UBE2L3 clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Ubiquitin Conjugating Enzyme E2L3 (UBE2L3). No significant cross-reactivity or interference between Ubiquitin Conjugating Enzyme E2L3 (UBE2L3) and analogues was observed.
Sequence Length	212
Assay Type	Double-antibody Sandwich
Samples	Tissue homogenates, Cell lysates and other biological fluids
Detection Range	13.7-10,000pg/mL
Sensitivity	5.7pg/mL
Application	Chemiluminescent immunoassay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Ubiquitin Conjugating Enzyme E2L3 (UBE2L3) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Ubiquitin Conjugating Enzyme E2L3 (UBE2L3) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for UBE2L3 clia kit	The microplate provided in this kit has been pre-coated with an antibody specific to Ubiquitin Conjugating Enzyme E2L3 (UBE2L3). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Ubiquitin Conjugating Enzyme E2L3 (UBE2L3). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Ubiquitin Conjugating Enzyme E2L3 (UBE2L3) level in the sample or standard.;
Product Categories/Family for UBE2L3 clia kit	Enzyme & Kinase

NCBI AND UNIPROT INFORMATION

NCBI GI #	373432684
NCBI GeneID	7332
NCBI Accession #	NP_001243284.1
NCBI GenBank Nucleotide #	NM_001256355.1
UniProt Accession #	P68036; P51966; P70653; Q9HAV1; B2R4A7; B4DDG1; B4DSZ4; E7EWS7
Molecular Weight	24,004 Da
NCBI Official Full Name	ubiquitin-conjugating enzyme E2 L3 isoform 4
NCBI Official Synonym Full Names	ubiquitin-conjugating enzyme E2L 3
NCBI Official Symbol	UBE2L3
NCBI Official Synonym Symbols	E2-F1; L-UBC; UBCH7; UbcM4
NCBI Protein Information	ubiquitin-conjugating enzyme E2 L3
UniProt Protein Name	Ubiquitin-conjugating enzyme E2 L3
UniProt Gene Name	UBE2L3
UniProt Synonym Gene Names	UBCE7; UBCH7
UniProt Entry Name	UB2L3_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.