

Human Tissue Factor (TF) CLIA Kit | TF clia kit

Human Tissue Factor (TF) CLIA Kit

CATALOG #
AAA130427

REACTIVITY
Human


PRODUCT SYNONYMS

Tissue Factor (TF); Human Tissue Factor (TF) CLIA Kit; CD142; FIII; F3; TFA; Thromboplastin; Coagulation Factor III; TF clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Tissue Factor (TF). No significant cross-reactivity or interference between Tissue Factor (TF) and analogues was observed.
Sequence Length	295
Assay Type	Double-antibody Sandwich
Samples	Plasma
Detection Range	2.74-2,000pg/mL
Sensitivity	0.97pg/mL
Application	Chemiluminescent immunoassay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Tissue Factor (TF) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Tissue Factor (TF) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for TF clia kit	The microplate provided in this kit has been pre-coated with an antibody specific to Tissue Factor (TF). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Tissue Factor (TF). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Tissue Factor (TF) level in the sample or standard.;
Product Categories/Family for TF clia kit	Hematology; Cardiovascular biology

NCBI AND UNIPROT INFORMATION

NCBI GI #	339508
NCBI GeneID	2152
UniProt Accession #	P13726; Q6FHG2; Q86WH4; D3DT47
Molecular Weight	27,145 Da
NCBI Official Full Name	tissue factor
NCBI Official Synonym Full Names	coagulation factor III (thromboplastin, tissue factor)
NCBI Official Symbol	F3
NCBI Official Synonym Symbols	TF; TFA; CD142
NCBI Protein Information	tissue factor
UniProt Protein Name	Tissue factor
UniProt Gene Name	F3
UniProt Synonym Gene Names	TF
UniProt Entry Name	TF_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.