

Rabbit Cholecystokinin A Receptor (CCKAR) CLIA Kit | CCKAR clia kit

Rabbit Cholecystokinin A Receptor (CCKAR) CLIA Kit

CATALOG #
AAA130568

REACTIVITY
Rabbit


PRODUCT SYNONYMS

Cholecystokinin A Receptor (CCKAR); Rabbit Cholecystokinin A Receptor (CCKAR) CLIA Kit; CCK-A; CCK1; CCKRA; CCK1R; Cholecystokinin Type 1 Receptor; CCKAR clia kit

PRODUCT SPECIFICATIONS

Reactivity	Rabbit
Specificity	This assay has high sensitivity and excellent specificity for detection of Cholecystokinin A Receptor (CCKAR). No significant cross-reactivity or interference between Cholecystokinin A Receptor (CCKAR) and analogues was observed.
Sequence Length	428
Assay Type	Double-antibody Sandwich
Samples	Tissue homogenates and other biological fluids
Detection Range	27.4-20,000pg/mL
Sensitivity	11.2pg/mL
Application	Chemiluminescent immunoassay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Cholecystokinin A Receptor (CCKAR) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Cholecystokinin A Receptor (CCKAR) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CCKAR clia kit	The microplate provided in this kit has been pre-coated with an antibody specific to Cholecystokinin A Receptor (CCKAR). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Cholecystokinin A Receptor (CCKAR). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Cholecystokinin A Receptor (CCKAR) level in the sample or standard.;
Product Categories/Family for CCKAR clia kit	Signal transduction; Endocrinology; Hepatology; Hormone metabolism

NCBI AND UNIPROT INFORMATION

NCBI GI #	4502607
NCBI GeneID	886
NCBI Accession #	NP_000721.1
NCBI GenBank Nucleotide #	NM_000730.2
UniProt Accession #	P32238; B2R9Z5
Molecular Weight	47,841 Da
NCBI Official Full Name	cholecystokinin receptor type A
NCBI Official Synonym Full Names	cholecystokinin A receptor
NCBI Official Symbol	CCKAR
NCBI Official Synonym Symbols	CCK-A; CCK1R; CCKRA; CCK1-R
NCBI Protein Information	cholecystokinin receptor type A
UniProt Protein Name	Cholecystokinin receptor type A
UniProt Gene Name	CCKAR
UniProt Synonym Gene Names	CCKRA; CCK-A receptor; CCK-AR; CCK1-R
UniProt Entry Name	CCKAR_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.