

Tissue Inhibitors Of Metalloproteinase 3 (TIMP3) CLIA Kit | TIMP3 clia kit

Rhesus monkey Tissue Inhibitors Of Metalloproteinase 3 (TIMP3) CLIA Kit

CATALOG # AAA130581	REACTIVITY Rhesus monkey
-------------------------------	------------------------------------


PRODUCT SYNONYMS

Tissue Inhibitors Of Metalloproteinase 3 (TIMP3); Rhesus monkey Tissue Inhibitors Of Metalloproteinase 3 (TIMP3) CLIA Kit; SFD; TIMP Metallopeptidase Inhibitor 3; Metallopeptidase Inhibitor 3; Sorsby Fundus Dystrophy; Pseudoinflammatory; TIMP3 clia kit

PRODUCT SPECIFICATIONS

Reactivity	Rhesus monkey
Specificity	This assay has high sensitivity and excellent specificity for detection of Tissue Inhibitors Of Metalloproteinase 3 (TIMP3). No significant cross-reactivity or interference between Tissue Inhibitors Of Metalloproteinase 3 (TIMP3) and analogues was observed.
Sequence Length	31
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma, Tissue homogenates, Cell culture supernates and other biological fluids
Detection Range	27.4-20,000pg/mL
Sensitivity	10.2pg/mL
Application	Chemiluminescent immunoassay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Tissue Inhibitors Of Metalloproteinase 3 (TIMP3) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Tissue Inhibitors Of Metalloproteinase 3 (TIMP3) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for TIMP3 clia kit	The microplate provided in this kit has been pre-coated with an antibody specific to Tissue Inhibitors Of Metalloproteinase 3 (TIMP3). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Tissue Inhibitors Of Metalloproteinase 3 (TIMP3). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Tissue Inhibitors Of Metalloproteinase 3 (TIMP3) level in the sample or standard.;
Product Categories/Family for TIMP3 clia kit	Tumor immunity; Cardiovascular biology

NCBI AND UNIPROT INFORMATION

NCBI GI #	16923457
Molecular Weight	3,604 Da
NCBI Official Full Name	tissue inhibitor of metalloproteinase 3, partial
UniProt Protein Name	Tissue inhibitor of metalloproteinase 3
UniProt Gene Name	TIMP3
UniProt Entry Name	Q8WP05_ATECH

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/monkey-rhesus-monkey-clia-kits/tissue-inhibitors-of-metalloproteinase-3-timp3/130581>