

Rat Endostatin (ES) ELISA Kit | ES elisa kit

Rat Endostatin (ES) ELISA Kit

CATALOG #
AAA135693

REACTIVITY
Rat


PRODUCT SYNONYMS

Endostatin (ES); Rat Endostatin (ES) ELISA Kit; ES elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This assay has high sensitivity and excellent specificity for detection of Endostatin (ES). No significant cross-reactivity or interference between Endostatin (ES) and analogues was observed.
Sequence Length	102
Samples	Serum, plasma and other biological fluids
Assay Type	Sandwich
Detection Range	31.25-2000pg/mL
Sensitivity	Typically less than 15.1pg/mL.
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Endostatin (ES) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Endostatin (ES) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Assay Procedure Summary	1. Prepare all reagents, samples and standards; 2. Add 100uL standard or sample to each well. Incubate 2 hours at 37 degree C; 3. Aspirate and add 100uL prepared Detection Reagent A. Incubate 1 hour at 37 degree C; 4. Aspirate and wash 3 times; 5. Add 100uL prepared Detection Reagent B. Incubate 30 minutes at 37 degree C; 6. Aspirate and wash 5 times; 7. Add 90uL Substrate Solution. Incubate 15-25 minutes at 37 degree C; 8. Add 50uL Stop Solution. Read at 450nm immediately.
Preparation and Storage	For unopened kit: All the reagents should be kept according to the labels on vials. The Standard, Detection Reagent A, Detection Reagent B and the 96-well strip plate should be stored at -20 degree C upon receipt while the others should be at 4 degree C. For opened kit: When the kit is opened, the remaining reagents still need to be stored according to the above storage condition. Besides, please return the unused wells to the foil pouch containing the desiccant pack, and reseal along entire edge of zip-seal. The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for ES elisa kit	Principle of the Assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Endostatin (ES). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Endostatin (ES). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Endostatin (ES), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Endostatin (ES) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
--	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	13957730
NCBI GeneID	80781
UniProt Accession #	P39060; Q58EX6; Q6RZ39; Q6RZ40; Q6RZ41; Q8N4S4; Q8WXI5; Q96T70; Q9UK38; Q9Y6Q7; Q9Y6Q8; A8MVI4
Molecular Weight	135,761 Da
NCBI Official Full Name	endostatin variant, partial
NCBI Official Synonym Full Names	collagen, type XVIII, alpha 1
NCBI Official Symbol	COL18A1
NCBI Official Synonym Symbols	KS; KNO; KNO1
NCBI Protein Information	collagen alpha-1(XVIII) chain
UniProt Protein Name	Collagen alpha-1(XVIII) chain
UniProt Gene Name	COL18A1
UniProt Entry Name	COIA1_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.