

Rat Potassium Inwardly Rectifying Channel Subfamily J, Member 5 (KCNJ5) ELISA Kit | KCNJ5 elisa kit

Rat Potassium Inwardly Rectifying Channel Subfamily J, Member 5 (KCNJ5) ELISA Kit

CATALOG # AAA136190	REACTIVITY Rat
-------------------------------	--------------------------


PRODUCT SYNONYMS

Potassium Inwardly Rectifying Channel Subfamily J; Member 5 (KCNJ5); Rat Potassium Inwardly Rectifying Channel Subfamily J; Member 5 (KCNJ5) ELISA Kit; Kir3.4; CIR; KATP1; GIRK4; G protein-activated inward rectifier potassium channel 4; Cardiac inward rectifier; Inward rectifier K(+) channel Kir3.4; Heart KATP channel; KCNJ5 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This assay has high sensitivity and excellent specificity for detection of Potassium Inwardly Rectifying Channel Subfamily J, Member 5 (KCNJ5). No significant cross-reactivity or interference between Potassium Inwardly Rectifying Channel Subfamily J, Member 5 (KCNJ5) and analogues was observed.
Sequence Length	404
Assay Type	Double-antibody Sandwich
Samples	Tissue Homogenates And Other Biological Fluids.
Detection Range	0.156-10ng/mL
Sensitivity	0.055ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Potassium Inwardly Rectifying Channel Subfamily J, Member 5 (KCNJ5) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Potassium Inwardly Rectifying Channel Subfamily J, Member 5 (KCNJ5) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for KCNJ5 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Potassium Inwardly Rectifying Channel Subfamily J, Member 5 (KCNJ5). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Potassium Inwardly Rectifying Channel Subfamily J, Member 5 (KCNJ5). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Potassium Inwardly Rectifying Channel Subfamily J, Member 5 (KCNJ5), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Potassium Inwardly Rectifying Channel Subfamily J, Member 5 (KCNJ5) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Product Categories/Family for KCNJ5 elisa kit	Signal Transduction

NCBI AND UNIPROT INFORMATION

NCBI GI #	62857935
NCBI GeneID	549655
NCBI Accession #	NP_001016901.1
NCBI GenBank Nucleotide #	NM_001016901.2
NCBI Official Full Name	G protein-activated inward rectifier potassium channel 4
NCBI Official Synonym Full Names	potassium channel, inwardly rectifying subfamily J, member 5
NCBI Official Symbol	kcnj5
NCBI Official Synonym Symbols	cir; girk4; katp1; kcnj6; kir3.4
NCBI Protein Information	G protein-activated inward rectifier potassium channel 4

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/rat-elisa-kits/potassium-inwardly-rectifying-channel-subfamily-j-member-5-kcnj5/136190>