

Human Flotillin 2 (FLOT2) ELISA Kit | FLOT2 elisa kit

Human Flotillin 2 (FLOT2) ELISA Kit

CATALOG #
AAA136385

REACTIVITY
Human


PRODUCT SYNONYMS

Flotillin 2 (FLOT2); Human Flotillin 2 (FLOT2) ELISA Kit; ECS1; ESA; ESA1; M17S1; Epidermal Surface Antigen 1; Membrane Component; Chromosome 17; Surface Marker 1; 35kD Protein Identified By Monoclonal Antibody ECS-1; FLOT2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Flotillin 2 (FLOT2). No significant cross-reactivity or interference between Flotillin 2 (FLOT2) and analogues was observed.
Sequence Length	438
Assay Type	Double-antibody Sandwich
Samples	Tissue homogenates, Cell lysates and other biological fluids
Detection Range	0.625-40ng/mL
Sensitivity	0.217ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Flotillin 2 (FLOT2) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Flotillin 2 (FLOT2) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for FLOT2 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Flotillin 2 (FLOT2). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Flotillin 2 (FLOT2). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Flotillin 2 (FLOT2), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Flotillin 2 (FLOT2) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Product Categories/Family for FLOT2 elisa kit	Neuroscience

NCBI AND UNIPROT INFORMATION

NCBI GI #	442616365
NCBI GeneID	32425
NCBI Accession #	NP_001259555.1
NCBI GenBank Nucleotide #	NM_001272626.2
UniProt Accession #	O61492; Q0KHS9; Q8SZW0; Q9I7S0; Q9VXZ9
Molecular Weight	31,008 Da
NCBI Official Full Name	flotillin 2, isoform K
NCBI Official Synonym Full Names	Flotillin 2
NCBI Official Symbol	Flo2
NCBI Official Synonym Symbols	CG11547; CG14409; CG32593; DmelCG32593; DmFlo2; Flo; Flo-2; flo2; FLO2; FLO[Dm]-2; reggie-1
NCBI Protein Information	CG32593 gene product from transcript CG32593-RJ
UniProt Protein Name	Flotillin-2
UniProt Gene Name	Flo-2
UniProt Synonym Gene Names	FLODm-2
UniProt Entry Name	FLOT2_DROME

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/flotillin-2-flot2/136385>