

Human Cryptochrome 1 (CRY1) ELISA Kit | CRY1 elisa kit

Human Cryptochrome 1 (CRY1) ELISA Kit

CATALOG #
AAA137059

REACTIVITY
Human


PRODUCT SYNONYMS

Cryptochrome 1 (CRY1); Human Cryptochrome 1 (CRY1) ELISA Kit; PHLL1; Photolyase-like 1; CRY1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Cryptochrome 1 (CRY1). No significant cross-reactivity or interference between Cryptochrome 1 (CRY1) and analogues was observed.
Sequence Length	681
Assay Type	Double-antibody Sandwich
Samples	Tissue homogenates, Cell lysates and other biological fluids
Detection Range	0.312-20ng/mL
Sensitivity	0.118ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Cryptochrome 1 (CRY1) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Cryptochrome 1 (CRY1) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CRY1 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Cryptochrome 1 (CRY1). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Cryptochrome 1 (CRY1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Cryptochrome 1 (CRY1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Cryptochrome 1 (CRY1) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Product Categories/Family for CRY1 elisa kit	Neuroscience; Nutrition metabolism; Circadian rhythm

NCBI AND UNIPROT INFORMATION

NCBI GI #	18413170
NCBI GeneID	826470
NCBI Accession #	NP_567341.1
NCBI GenBank Nucleotide #	NM_116961.4
UniProt Accession #	Q43125; Q43126; Q8L7Y1; Q9ASZ2; Q9M0S9; Q9ZPF0
Molecular Weight	76,695 Da
NCBI Official Full Name	cryptochrome-1
NCBI Official Symbol	CRY1
NCBI Official Synonym Symbols	ATCRY1; BLU1; BLUE LIGHT UNINHIBITED 1; CRY1; cryptochrome 1; CRYPTOCHROME 1 APOPROTEIN (BLUE LIGHT PHOTORECEPTOR; ELONGATED HYPOCOTYL 4; HY4; OOP2; OUT OF PHASE 2; T3H13.14; T3H13_14
NCBI Protein Information	cryptochrome-1
UniProt Protein Name	Cryptochrome-1
UniProt Gene Name	CRY1
UniProt Synonym Gene Names	BLU1; HY4; OOP2
UniProt Entry Name	CRY1_ARATH

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/cryptochrome-1-cry1/137059>