

Human Aquaporin 9 (AQP9) ELISA Kit | AQP9 elisa kit

Human Aquaporin 9 (AQP9) ELISA Kit

CATALOG #
AAA137186

REACTIVITY
Human


PRODUCT SYNONYMS

Aquaporin 9 (AQP9); Human Aquaporin 9 (AQP9) ELISA Kit; SSC1; Aquaglyceroporin-9; Small solute channel 1; AQP9 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Aquaporin 9 (AQP9). No significant cross-reactivity or interference between Aquaporin 9 (AQP9) and analogues was observed.
Sequence Length	24
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma, Tissue homogenates and other biological fluids
Detection Range	15.6-1,000pg/mL
Sensitivity	6.4pg/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Aquaporin 9 (AQP9) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Aquaporin 9 (AQP9) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for AQP9 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Aquaporin 9 (AQP9). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Aquaporin 9 (AQP9). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Aquaporin 9 (AQP9), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Aquaporin 9 (AQP9) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Product Categories/Family for AQP9 elisa kit	Metabolic pathway; Infection immunity; Developmental science

NCBI AND UNIPROT INFORMATION

NCBI GI #	6759671
NCBI GeneID	366
UniProt Accession #	O43315; Q9NP32
Molecular Weight	31,431 Da
NCBI Official Full Name	aquaporin 9, partial
NCBI Official Synonym Full Names	aquaporin 9
NCBI Official Symbol	AQP9
NCBI Official Synonym Symbols	SSC1; AQP-9; HsT17287
NCBI Protein Information	aquaporin-9
UniProt Protein Name	Aquaporin-9
UniProt Gene Name	AQP9
UniProt Synonym Gene Names	SSC1; AQP-9
UniProt Entry Name	AQP9_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/aquaporin-9-aqp9/137186>