

Human Mannose Associated Serine Protease 2 (MASP2) ELISA Kit | MASP2 elisa kit

Human Mannose Associated Serine Protease 2 (MASP2) ELISA Kit

CATALOG #
AAA137840


REACTIVITY
Human

PRODUCT SYNONYMS

Mannose Associated Serine Protease 2 (MASP2); Human Mannose Associated Serine Protease 2 (MASP2) ELISA Kit; Mannose-binding protein-associated serine protease 2; mannan-Binding Lectin Serine Peptidase 2; MBL-associated serine protease 2; MASP2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Mannose Associated Serine Protease 2 (MASP2). No significant cross-reactivity or interference between Mannose Associated Serine Protease 2 (MASP2) and analogues was observed.
Sequence Length	615
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma and other biological fluids
Detection Range	31.2-2,000pg/mL
Sensitivity	12.4pg/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Mannose Associated Serine Protease 2 (MASP2) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Mannose Associated Serine Protease 2 (MASP2) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for MASP2 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Mannose Associated Serine Protease 2 (MASP2). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Mannose Associated Serine Protease 2 (MASP2). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Mannose Associated Serine Protease 2 (MASP2), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Mannose Associated Serine Protease 2 (MASP2) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Product Categories/Family for MASP2 elisa kit	Enzyme & Kinase; Infection immunity; Immune molecule

NCBI AND UNIPROT INFORMATION

NCBI GI #	21264359
NCBI GeneID	5648
NCBI Accession #	NP_624302.1
NCBI GenBank Nucleotide #	NM_139125.3
UniProt Accession #	P48740; O95570; Q68D21; Q8IUV8; Q96RS4; Q9UF09; A8K542; A8K6M1; B4E2L7
Molecular Weight	68,918 Da
NCBI Official Full Name	mannan-binding lectin serine protease 1 isoform 2
NCBI Official Synonym Full Names	mannan-binding lectin serine peptidase 1 (C4/C2 activating component of Ra-reactive factor)
NCBI Official Symbol	MASP1
NCBI Official Synonym Symbols	3MC1; MAP1; MASP; RaRF; CRARF; MASP3; Map44; PRSS5; CRARF1
NCBI Protein Information	mannan-binding lectin serine protease 1
UniProt Protein Name	Mannan-binding lectin serine protease 1
UniProt Gene Name	MASP1
UniProt Synonym Gene Names	CRARF; CRARF1; PRSS5; MASP-1; RaRF
UniProt Entry Name	MASP1_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.