

Human Endoplasmic Reticulum Aminopeptidase 2 (ERAP2) ELISA Kit | ERAP2 elisa kit

Human Endoplasmic Reticulum Aminopeptidase 2 (ERAP2) ELISA Kit

CATALOG #
AAA138161

REACTIVITY
Human


PRODUCT SYNONYMS

Endoplasmic Reticulum Aminopeptidase 2 (ERAP2); Human Endoplasmic Reticulum Aminopeptidase 2 (ERAP2) ELISA Kit; L-RAP; LRAP; Leukocyte-Derived Arginine Aminopeptidase; ERAP2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Endoplasmic Reticulum Aminopeptidase 2 (ERAP2). No significant cross-reactivity or interference between Endoplasmic Reticulum Aminopeptidase 2 (ERAP2) and analogues was observed.
Sequence Length	350
Assay Type	Double-antibody Sandwich
Samples	Tissue homogenates, Cell lysates and other biological fluids
Detection Range	0.312-20ng/mL
Sensitivity	0.115ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Endoplasmic Reticulum Aminopeptidase 2 (ERAP2) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Endoplasmic Reticulum Aminopeptidase 2 (ERAP2) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for ERAP2 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Endoplasmic Reticulum Aminopeptidase 2 (ERAP2). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Endoplasmic Reticulum Aminopeptidase 2 (ERAP2). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Endoplasmic Reticulum Aminopeptidase 2 (ERAP2), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Endoplasmic Reticulum Aminopeptidase 2 (ERAP2) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Product Categories/Family for ERAP2 elisa kit	Enzyme & Kinase

NCBI AND UNIPROT INFORMATION

NCBI GI #	11641261
NCBI GeneID	64167
NCBI Accession #	NP_071745.1
NCBI GenBank Nucleotide #	NM_022350.3
UniProt Accession #	Q6P179; Q7Z5K1; Q8TD32; Q8WVJ4; Q9HBX2
Molecular Weight	40,088 Da
NCBI Official Full Name	endoplasmic reticulum aminopeptidase 2
NCBI Official Synonym Full Names	endoplasmic reticulum aminopeptidase 2
NCBI Official Symbol	ERAP2
NCBI Official Synonym Symbols	LRAP; L-RAP
NCBI Protein Information	endoplasmic reticulum aminopeptidase 2
UniProt Protein Name	Endoplasmic reticulum aminopeptidase 2
UniProt Gene Name	ERAP2
UniProt Synonym Gene Names	LRAP; L-RAP
UniProt Entry Name	ERAP2_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.