

Mouse Phosphoinositide-3-Kinase Class-2-Alpha Polypeptide (PIK3C2a) ELISA Kit | PIK3C2a elisa kit

Mouse Phosphoinositide-3-Kinase Class-2-Alpha Polypeptide (PIK3C2a) ELISA Kit

CATALOG #
AAA138292

REACTIVITY
Mouse


PRODUCT SYNONYMS

Phosphoinositide-3-Kinase Class-2-Alpha Polypeptide (PIK3C2a); Mouse Phosphoinositide-3-Kinase Class-2-Alpha Polypeptide (PIK3C2a) ELISA Kit; PI3K2a; CPK; PI3-K-C2(Alpha); PI3-K-C2A; PI3K2-a; Phosphoinositide 3-Kinase 2a; Phosphatidylinositol 4-phosphate 3-kinase C2 domain-containing subunit alpha; PIK3C2a elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Mouse
Specificity	This assay has high sensitivity and excellent specificity for detection of Phosphoinositide-3-Kinase Class-2-Alpha Polypeptide (PIK3C2a). No significant cross-reactivity or interference between Phosphoinositide-3-Kinase Class-2-Alpha Polypeptide (PIK3C2a) and analogues was observed.
Sequence Length	483
Assay Type	Double-antibody Sandwich
Samples	Tissue homogenates and other biological fluids
Detection Range	0.156-10ng/mL
Sensitivity	0.057ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Phosphoinositide-3-Kinase Class-2-Alpha Polypeptide (PIK3C2a) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Phosphoinositide-3-Kinase Class-2-Alpha Polypeptide (PIK3C2a) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for PIK3C2a elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Phosphoinositide-3-Kinase Class-2-Alpha Polypeptide (PIK3C2a). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Phosphoinositide-3-Kinase Class-2-Alpha Polypeptide (PIK3C2a). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Phosphoinositide-3-Kinase Class-2-Alpha Polypeptide (PIK3C2a), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Phosphoinositide-3-Kinase Class-2-Alpha Polypeptide (PIK3C2a) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Product Categories/Family for PIK3C2a elisa kit	Signal Transduction

NCBI AND UNIPROT INFORMATION

NCBI GI #	119588853
NCBI GeneID	5286
Molecular Weight	53,742 Da
NCBI Official Full Name	phosphoinositide-3-kinase, class 2, alpha polypeptide
NCBI Official Synonym Full Names	phosphatidylinositol-4-phosphate 3-kinase, catalytic subunit type 2 alpha
NCBI Official Symbol	PIK3C2A
NCBI Official Synonym Symbols	CPK; PI3-K-C2A; PI3-K-C2(ALPHA)
NCBI Protein Information	phosphatidylinositol 4-phosphate 3-kinase C2 domain-containing subunit alpha
UniProt Protein Name	Phosphatidylinositol 4-phosphate 3-kinase C2 domain-containing subunit alpha
UniProt Gene Name	PIK3C2A
UniProt Synonym Gene Names	PI3K-C2-alpha; PtdIns-3-kinase C2 subunit alpha
UniProt Entry Name	P3C2A_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.