

Human Adenosine A2a Receptor (ADORA2a) ELISA Kit | ADORA2a elisa kit

Human Adenosine A2a Receptor (ADORA2a) ELISA Kit

CATALOG #
AAA138905

REACTIVITY
Human


PRODUCT SYNONYMS

Adenosine A2a Receptor (ADORA2a); Human Adenosine A2a Receptor (ADORA2a) ELISA Kit; ADORA2; RDC8; hA2AR; ADORA2a elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Adenosine A2a Receptor (ADORA2a). No significant cross-reactivity or interference between Adenosine A2a Receptor (ADORA2a) and analogues was observed.
Sequence Length	412
Assay Type	Double-antibody Sandwich
Samples	Tissue homogenates and other biological fluids
Detection Range	0.156-10ng/mL
Sensitivity	0.060ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Adenosine A2a Receptor (ADORA2a) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Adenosine A2a Receptor (ADORA2a) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for ADORA2a elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Adenosine A2a Receptor (ADORA2a). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Adenosine A2a Receptor (ADORA2a). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Adenosine A2a Receptor (ADORA2a), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Adenosine A2a Receptor (ADORA2a) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Product Categories/Family for ADORA2a elisa kit	CD & Adhesion molecule

NCBI AND UNIPROT INFORMATION

NCBI GI #	119580062
NCBI GeneID	135
UniProt Accession #	P29274; B2R7E0
Molecular Weight	44,707 Da
NCBI Official Full Name	adenosine A2a receptor
NCBI Official Synonym Full Names	adenosine A2a receptor
NCBI Official Symbol	ADORA2A
NCBI Official Synonym Symbols	A2aR; RDC8; ADORA2
NCBI Protein Information	adenosine receptor A2a
UniProt Protein Name	Adenosine receptor A2a
UniProt Gene Name	ADORA2A
UniProt Synonym Gene Names	ADORA2
UniProt Entry Name	AA2AR_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/adenosine-a2a-receptor-adora2a/138905>