

Human Dopa Decarboxylase (DDC) ELISA Kit | DDC elisa kit

Human Dopa Decarboxylase (DDC) ELISA Kit

CATALOG #
AAA138973

REACTIVITY
Human


PRODUCT SYNONYMS

Dopa Decarboxylase (DDC); Human Dopa Decarboxylase (DDC) ELISA Kit; AADC; Aromatic L-Amino Acid Decarboxylase; DOPA decarboxylase; DDC elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of DDC. No significant cross-reactivity or interference between DDC and analogues was observed.
Sequence Length	475
Assay Type	Double-antibody Sandwich
Samples	Tissue homogenates, Cell lysates and other biological fluids
Detection Range	1.56-100ng/mL
Sensitivity	0.61ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level DDC were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level DDC were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for DDC elisa kit	Intended Uses: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of DDC in human tissue homogenates, cell lysates and other biological fluids. Principle of the Assay: The microplate provided in this kit has been pre-coated with an antibody specific to DDC. Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to DDC. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain DDC, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm +/- 10nm. The concentration of DDC in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Product Categories/Family for DDC elisa kit	Enzyme & Kinase

NCBI AND UNIPROT INFORMATION

NCBI GI #	221476190
NCBI GeneID	35190
NCBI Accession #	NP_523600.5
NCBI GenBank Nucleotide #	NM_078876.5
Molecular Weight	56,863 Da
NCBI Official Full Name	dopa decarboxylase, isoform D
NCBI Official Synonym Full Names	Dopa decarboxylase
NCBI Official Symbol	Ddc
NCBI Official Synonym Symbols	5-HT; AADC; CG10697; ddc; DDC; DdcDm; DmelCG10697; fDDC; l(2)37Bl; l(2)37Ch; l(2)k02104
NCBI Protein Information	CG10697 gene product from transcript CG10697-RB
UniProt Protein Name	Aromatic-L-amino-acid decarboxylase
UniProt Gene Name	Ddc
UniProt Synonym Gene Names	AADC; DDC
UniProt Entry Name	DDC_DROME

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.