

Human Guanylate Cyclase Activator 2A (GUCA2A) ELISA Kit | GUCA2A elisa kit

Human Guanylate Cyclase Activator 2A (GUCA2A) ELISA Kit

CATALOG #
AAA139244


REACTIVITY
Human

PRODUCT SYNONYMS

Guanylate Cyclase Activator 2A (GUCA2A); Human Guanylate Cyclase Activator 2A (GUCA2A) ELISA Kit; GUCA2; STARA; GCAP-I; Guanylin; Guanylate cyclase-activating protein 1; GUCA2A elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Guanylate Cyclase Activator 2A (GUCA2A). No significant cross-reactivity or interference between Guanylate Cyclase Activator 2A (GUCA2A) and analogues was observed.
Samples	Serum, plasma and other biological fluids
Assay Type	Competitive
Detection Range	24.69-2000fmol/mL
Sensitivity	Typically less than 8.78fmol/mL.
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Guanylate Cyclase Activator 2A (GUCA2A) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Guanylate Cyclase Activator 2A (GUCA2A) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Assay Procedure Summary	1. Prepare all reagents, samples and standards; 2. Add 50uL standard or sample to each well. And then add 50uL prepared Detection Reagent A immediately. Shake and mix. Incubate 1 hour at 37 degree C; 3. Aspirate and wash 3 times; 4. Add 100uL prepared Detection Reagent B. Incubate 30 minutes at 37 degree C; 5. Aspirate and wash 5 times; 6. Add 90uL Substrate Solution. Incubate 15-25 minutes at 37 degree C; 7. Add 50uL Stop Solution. Read at 450 nm immediately.
Preparation and Storage	For unopened kit: All the reagents should be kept according to the labels on vials. The Standard, Detection Reagent A, Detection Reagent B and the 96-well strip plate should be stored at -20 degree C upon receipt while the others should be at 4 degree C. For opened kit: When the kit is opened, the remaining reagents still need to be stored according to the above storage condition. Besides, please return the unused wells to the foil pouch containing the desiccant pack, and reseal along entire edge of zip-seal. The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for
GUCA2A elisa kit

Principle of the Assay: This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to Guanylate Cyclase Activator 2A (GUCA2A) has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled Guanylate Cyclase Activator 2A (GUCA2A) and unlabeled Guanylate Cyclase Activator 2A (GUCA2A) (Standards or samples) with the pre-coated antibody specific to Guanylate Cyclase Activator 2A (GUCA2A). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Guanylate Cyclase Activator 2A (GUCA2A) in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of Guanylate Cyclase Activator 2A (GUCA2A) in the sample.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/guanylate-cyclase-activator-2a-guca2a/139244>