

Human GM2 Ganglioside Activator (GM2A) ELISA Kit | GM2A elisa kit

Human GM2 Ganglioside Activator (GM2A) ELISA Kit

CATALOG #
AAA139425

REACTIVITY
Human


PRODUCT SYNONYMS

GM2 Ganglioside Activator (GM2A); Human GM2 Ganglioside Activator (GM2A) ELISA Kit; SAP-3; Cerebroside Sulfate Activator Protein; Sphingolipid Activator Protein 3; GM2A elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of GM2 Ganglioside Activator (GM2A). No significant cross-reactivity or interference between GM2 Ganglioside Activator (GM2A) and analogues was observed.
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma, Tissue homogenates, Cell lysates, Cell culture supernates and other biological fluids
Detection Range	0.156-10ng/mL
Sensitivity	0.064ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level GM2 Ganglioside Activator (GM2A) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level GM2 Ganglioside Activator (GM2A) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for GM2A elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to GM2 Ganglioside Activator (GM2A). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to GM2 Ganglioside Activator (GM2A). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain GM2 Ganglioside Activator (GM2A), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of GM2 Ganglioside Activator (GM2A) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Product Categories/Family for GM2A elisa kit	Immune molecule

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/gm2-ganglioside-activator-gm2a/139425>