

Human Signal Transducer And Activator Of Transcription 1 (STAT1) ELISA Kit | STAT1 elisa kit

Human Signal Transducer And Activator Of Transcription 1 (STAT1) ELISA Kit

CATALOG # AAA140336	REACTIVITY Human
-------------------------------	----------------------------


PRODUCT SYNONYMS

Signal Transducer And Activator Of Transcription 1 (STAT1); Human Signal Transducer And Activator Of Transcription 1 (STAT1) ELISA Kit; STAT1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Signal Transducer And Activator Of Transcription 1 (STAT1). No significant cross-reactivity or interference between Signal Transducer And Activator Of Transcription 1 (STAT1) and analogues was observed.
Sequence Length	750
Assay Type	Double-antibody Sandwich
Samples	Tissue homogenates, Cell lysates and other biological fluids
Detection Range	0.156-10ng/mL
Sensitivity	0.057ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Signal Transducer And Activator Of Transcription 1 (STAT1) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Signal Transducer And Activator Of Transcription 1 (STAT1) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for STAT1 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Signal Transducer And Activator Of Transcription 1 (STAT1). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Signal Transducer And Activator Of Transcription 1 (STAT1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Signal Transducer And Activator Of Transcription 1 (STAT1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Signal Transducer And Activator Of Transcription 1 (STAT1) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Product Categories/Family for STAT1 elisa kit	Signal transduction; Tumor immunity; Infection immunity

NCBI AND UNIPROT INFORMATION

NCBI GI #	281313277
NCBI GeneID	6772
UniProt Accession #	P42224; Q53S88; Q53XW4; Q68D00; Q9UDL5; A8K989; B2RCA0; D2KFR8; D3DPI7
Molecular Weight	87,335 Da
NCBI Official Full Name	signal transducer and activator of transcription 1
NCBI Official Synonym Full Names	signal transducer and activator of transcription 1, 91kDa
NCBI Official Symbol	STAT1
NCBI Official Synonym Symbols	CANDF7; ISGF-3; STAT91
NCBI Protein Information	signal transducer and activator of transcription 1-alpha/beta; transcription factor ISGF-3 components p91/p84; signal transducer and activator of transcription-1
UniProt Protein Name	Signal transducer and activator of transcription 1-alpha/beta
UniProt Gene Name	STAT1
UniProt Entry Name	STAT1_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/signal-transducer-and-activator-of-transcription-1-stat1/140336>