

Human Plasminogen Activator, Urokinase (uPA) CLIA Kit | uPA clia kit

Human Plasminogen Activator, Urokinase (uPA) CLIA Kit

CATALOG #
AAA140970

REACTIVITY
Human


PRODUCT SYNONYMS

Plasminogen Activator; Urokinase (uPA); Human Plasminogen Activator; Urokinase (uPA) CLIA Kit; PLAU; ATF; URK; UK; UP-A; Abbokinase; Urokinase-Type Plasminogen Activator; uPA clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Plasminogen Activator, Urokinase (uPA). No significant cross-reactivity or interference between Plasminogen Activator, Urokinase (uPA) and analogues was observed.
Sequence Length	431
Assay Type	Double-antibody Sandwich
Samples	Plasma, Tissue homogenates, Cell lysates, Cell culture supernates and other biological fluids
Detection Range	2.74-2,000pg/mL
Sensitivity	1.24pg/mL
Application	Chemiluminescent immunoassay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Plasminogen Activator, Urokinase (uPA) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Plasminogen Activator, Urokinase (uPA) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for uPA clia kit	The microplate provided in this kit has been pre-coated with an antibody specific to Plasminogen Activator, Urokinase (uPA). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Plasminogen Activator, Urokinase (uPA). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Plasminogen Activator, Urokinase (uPA) level in the sample or standard.;
Product Categories/Family for uPA clia kit	Tumor immunity

NCBI AND UNIPROT INFORMATION

NCBI GI #	15488889
NCBI GeneID	5328
Molecular Weight	46,908 Da
NCBI Official Full Name	Plasminogen activator, urokinase
NCBI Official Synonym Full Names	plasminogen activator, urokinase
NCBI Official Symbol	PLAU
NCBI Official Synonym Symbols	ATF; QPD; UPA; URK; u-PA; BDPLT5
NCBI Protein Information	urokinase-type plasminogen activator
UniProt Protein Name	Urokinase-type plasminogen activator
UniProt Gene Name	PLAU
UniProt Synonym Gene Names	U-plasminogen activator; uPA
UniProt Entry Name	UROK_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-clia-kits/plasminogen-activator-urokinase-upa/140970>