

Human Elastin Microfibril Interface Located Protein 2 (EMILIN2) CLIA Kit | EMILIN2 clia kit

Human Elastin Microfibril Interface Located Protein 2 (EMILIN2) CLIA Kit

CATALOG #
AAA141006

REACTIVITY
Human


PRODUCT SYNONYMS

Elastin Microfibril Interface Located Protein 2 (EMILIN2); Human Elastin Microfibril Interface Located Protein 2 (EMILIN2) CLIA Kit; FOAP-10; Elastin Microfibril Interfacer 2; EMILIN2 clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Elastin Microfibril Interface Located Protein 2 (EMILIN2). No significant cross-reactivity or interference between Elastin Microfibril Interface Located Protein 2 (EMILIN2) and analogues was observed.
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma, Tissue homogenates and other biological fluids
Detection Range	27.4-20,000pg/mL
Sensitivity	8.9pg/mL
Application	Chemiluminescent immunoassay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Elastin Microfibril Interface Located Protein 2 (EMILIN2) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Elastin Microfibril Interface Located Protein 2 (EMILIN2) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for EMILIN2 clia kit	The microplate provided in this kit has been pre-coated with an antibody specific to Elastin Microfibril Interface Located Protein 2 (EMILIN2). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Elastin Microfibril Interface Located Protein 2 (EMILIN2). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Elastin Microfibril Interface Located Protein 2 (EMILIN2) level in the sample or standard.;
Product Categories/Family for EMILIN2 clia kit	CD & Adhesion molecule

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-clia-kits/elastin-microfibril-interface-located-protein-2-emilin2/141006>