

Human Tissue Polypeptide Specific Antigen (TPS) CLIA Kit | TPS clia kit

Human Tissue Polypeptide Specific Antigen (TPS) CLIA Kit

CATALOG #
AAA141127

REACTIVITY
Human


PRODUCT SYNONYMS

Tissue Polypeptide Specific Antigen (TPS); Human Tissue Polypeptide Specific Antigen (TPS) CLIA Kit; Cytokine 18; Specific M3 Epitope Of TPA; TPS clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Tissue Polypeptide Specific Antigen (TPS). No significant cross-reactivity or interference between Tissue Polypeptide Specific Antigen (TPS) and analogues was observed.
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma and other biological fluids
Detection Range	27.4-20,000pg/mL
Sensitivity	10.5pg/mL
Application	Chemiluminescent immunoassay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Tissue Polypeptide Specific Antigen (TPS) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Tissue Polypeptide Specific Antigen (TPS) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for TPS clia kit	The microplate provided in this kit has been pre-coated with an antibody specific to Tissue Polypeptide Specific Antigen (TPS). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Tissue Polypeptide Specific Antigen (TPS). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Tissue Polypeptide Specific Antigen (TPS) level in the sample or standard.;
--	---

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-clia-kits/tissue-polypeptide-specific-antigen-tps/141127>