

Human Cytochrome C Oxidase Subunit I (COX1) ELISA Kit | COX1 elisa kit

Human Cytochrome C Oxidase Subunit I (COX1) ELISA Kit

CATALOG #
AAA141245

REACTIVITY
Human


PRODUCT SYNONYMS

Cytochrome C Oxidase Subunit I (COX1); Human Cytochrome C Oxidase Subunit I (COX1) ELISA Kit; MT-CO1; MTCO1; COI; COXI; Mitochondrially Encoded Cytochrome C Oxidase I; Cytochrome c oxidase polypeptide I; COX1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Cytochrome C Oxidase Subunit I (COX1). No significant cross-reactivity or interference between Cytochrome C Oxidase Subunit I (COX1) and analogues was observed.
Sequence Length	514
Assay Type	Double-antibody Sandwich
Samples	Tissue homogenates, Cell lysates and other biological fluids
Detection Range	0.312-20ng/mL
Sensitivity	0.114ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Cytochrome C Oxidase Subunit I (COX1) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Cytochrome C Oxidase Subunit I (COX1) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for COX1 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Cytochrome C Oxidase Subunit I (COX1). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Cytochrome C Oxidase Subunit I (COX1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Cytochrome C Oxidase Subunit I (COX1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Cytochrome C Oxidase Subunit I (COX1) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Product Categories/Family for COX1 elisa kit	Enzyme & Kinase

NCBI AND UNIPROT INFORMATION

NCBI GI #	34538600
NCBI GeneID	17708
NCBI Accession #	NP_904330.1
NCBI GenBank Nucleotide #	NC_005089.1
Molecular Weight	56,910 Da
NCBI Official Full Name	cytochrome c oxidase subunit I (mitochondrion)
NCBI Official Synonym Full Names	cytochrome c oxidase I, mitochondrial
NCBI Official Symbol	mt-Co1
NCBI Official Synonym Symbols	CoxI; COX1
NCBI Protein Information	cytochrome c oxidase subunit I
UniProt Protein Name	Cytochrome c oxidase subunit 1
UniProt Gene Name	mt-Co1
UniProt Entry Name	Q9MD68_MOUSE

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/cytochrome-c-oxidase-subunit-i-cox1/141245>