

Mouse Glycated Hemoglobin A1c (HbA1c) ELISA Kit | HbA1c elisa kit

Mouse Glycated Hemoglobin A1c (HbA1c) Mini Samples ELISA Kit

CATALOG #
AAA147326

REACTIVITY
Mouse


PRODUCT SYNONYMS

Glycated Hemoglobin A1c (HbA1c); Mouse Glycated Hemoglobin A1c (HbA1c) Mini Samples ELISA Kit; Glycosylated Hemoglobin; Hemoglobin A1c; Hb1c; HbA1c; HbA1c elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Mouse
Specificity	This assay has high sensitivity and excellent specificity for detection of HbA1c. No significant cross-reactivity or interference between HbA1c and analogues was observed.
Assay Type	Competitive Inhibition
Samples	Plasma and Erythrocyte lysates
Detection Range	12.35-1,000ug/mL
Sensitivity	4.94ug/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level HbA1c were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level HbA1c were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for HbA1c elisa kit

Intended Uses: The kit is a competitive inhibition enzyme immunoassay technique for the in vitro quantitative measurement of HbA1c in mouse plasma and erythrocyte lysates.

Principle of the Assay: This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to HbA1c has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled HbA1c and unlabeled HbA1c (Standards or samples) with the pre-coated antibody specific to HbA1c. After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of HbA1c in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of HbA1c in the sample.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/mouse-elisa-kits/glycated-hemoglobin-a1c-hba1c/147326>