

Human Glycated Hemoglobin A1c (HbA1c) CLIA Kit | HbA1c clia kit

Human Glycated Hemoglobin A1c (HbA1c) CLIA Kit

CATALOG #
AAA148378

REACTIVITY
Human


PRODUCT SYNONYMS

Glycated Hemoglobin A1c (HbA1c); Human Glycated Hemoglobin A1c (HbA1c) CLIA Kit; Glycosylated Hemoglobin; Hemoglobin A1c; Hb1c; HbA1c; HbA1c clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Glycated Hemoglobin A1c (HbA1c). No significant cross-reactivity or interference between Glycated Hemoglobin A1c (HbA1c) and analogues was observed.
Assay Type	Competitive Inhibition
Samples	Serum, Plasma and Erythrocyte lysates
Detection Range	5.86-1,500ug/mL
Sensitivity	2.44ug/mL
Application	Chemiluminescent immunoassay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Glycated Hemoglobin A1c (HbA1c) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Glycated Hemoglobin A1c (HbA1c) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for HbA1c clia kit	The microplate provided in this kit has been pre-coated with a monoclonal antibody specific to Glycated Hemoglobin A1c (HbA1c). A competitive inhibition reaction is launched between biotin labeled Glycated Hemoglobin A1c (HbA1c) and unlabeled Glycated Hemoglobin A1c (HbA1c) (Standards or samples) with the pre-coated antibody specific to Glycated Hemoglobin A1c (HbA1c). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Glycated Hemoglobin A1c (HbA1c) in the sample. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is reverse proportional to the Glycated Hemoglobin A1c (HbA1c) level in the sample or standard.
--	--

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-clia-kits/glycated-hemoglobin-a1c-hba1c/148378>