

## Human Complement 3, C3 ELISA Kit | C3 elisa kit

Human Complement 3, C3 ELISA Kit

CATALOG #  
**AAA15098**

REACTIVITY  
**Human**


### PRODUCT SYNONYMS

Complement 3; C3; Human Complement 3; C3 ELISA Kit; C3 elisa kit

### PRODUCT SPECIFICATIONS

|                         |                                                                                                                                                                                                                                                                                                                                                           |
|-------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Reactivity              | Human                                                                                                                                                                                                                                                                                                                                                     |
| Specificity             | This assay has high sensitivity and excellent specificity for detection of human C3. No significant cross-reactivity or interference between human C3 and analogues was observed.                                                                                                                                                                         |
| Samples                 | Serum, plasma, cell culture supernates, urine, saliva                                                                                                                                                                                                                                                                                                     |
| Assay Type              | Sandwich                                                                                                                                                                                                                                                                                                                                                  |
| Detection Range         | 4.7 ng/ml -300 ng/ml.                                                                                                                                                                                                                                                                                                                                     |
| Sensitivity             | The minimum detectable dose of human C3 is typically less than 4.031 ng/ml. The sensitivity of this assay, or Lower Limit of Detection (LLD) was defined as the lowest protein concentration that could be differentiated from zero. It was determined the mean O.D value of 20 replicates of the zero standard added by their three standard deviations. |
| Intra-assay Precision   | Intra-assay Precision (Precision within an assay): CV%<8%. Three samples of known concentration were tested twenty times on one plate to assess.                                                                                                                                                                                                          |
| Inter-assay Precision   | Inter-assay Precision (Precision between assays): CV%<10%. Three samples of known concentration were tested in twenty assays to assess.                                                                                                                                                                                                                   |
| Preparation and Storage | Unopened test kits should be stored at 2 to 8 degree C upon receipt. Please refer to pdf manual for further storage instructions.                                                                                                                                                                                                                         |

### IMAGES



Standard Curve (Sample)

### ADDITIONAL INFORMATION

#### Related Product Information for C3 elisa kit

Principle of the assay: This assay employs the quantitative sandwich enzyme immunoassay technique. Antibody specific for C3 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any C3 present is bound by the immobilized antibody. After removing any unbound substances, a Horseradish Peroxidase (HRP)-conjugated antibody specific for C3 is added to the wells. Following a wash to remove any unbound reagent, a substrate solution is added to the wells and color develops in proportion to the amount of C3 bound in the initial step. The color development is stopped and the intensity of the color is measured.

**For Research Use Only.** Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | [contact@aaabiotech.com](mailto:contact@aaabiotech.com)  
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/complement-3-c3/15098>