

Human Receptor Tyrosine Kinase Like Orphan Receptor 1 (ROR1) CLIA Kit | ROR1 clia kit

Human Receptor Tyrosine Kinase Like Orphan Receptor 1 (ROR1)(Sandwich CLIA) CLIA Kit

CATALOG #
AAA152472

REACTIVITY
Human


PRODUCT SYNONYMS

Receptor Tyrosine Kinase Like Orphan Receptor 1 (ROR1); Human Receptor Tyrosine Kinase Like Orphan Receptor 1 (ROR1)(Sandwich CLIA) CLIA Kit; CLIA Kit for Receptor Tyrosine Kinase Like Orphan Receptor 1 (ROR1); NTRKR1; Tyrosine-protein kinase transmembrane receptor ROR1; Neurotrophic tyrosine kinase; receptor-related 1; ROR1 clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of ROR1. No significant cross-reactivity or interference between ROR1 and analogues was observed.
Samples	Human serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Detection Range	68.6-50,000pg/mL
Sensitivity	< 23.1pg/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level ROR1 were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level ROR1 were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	Store at 4 degree C for 1 month/-20 degree C for long-term (within shelf life) Shelf-life: one year The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition.To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for ROR1 clia kit	Principle of the Assay: The microplate provided in this kit has been pre-coated with an antibody specific to ROR1. Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to ROR1. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the ROR1 level in the sample or standard. Intended Uses: The Chemiluminescent Immunoassay kit is designed for the in vitro sensitive quantitative measurement of ROR1 in human serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids.
---	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	118572711
NCBI GeneID	4919
UniProt Accession #	Q01973; Q5VVX6; Q66K77; Q92776
Molecular Weight	937
NCBI Official Full Name	Tyrosine-protein kinase transmembrane receptor ROR1
NCBI Official Synonym Full Names	receptor tyrosine kinase-like orphan receptor 1
NCBI Official Symbol	ROR1
NCBI Official Synonym Symbols	NTRKR1; dJ537F10.1
NCBI Protein Information	tyrosine-protein kinase transmembrane receptor ROR1; neurotrophic tyrosine kinase, receptor-related 1
UniProt Protein Name	Tyrosine-protein kinase transmembrane receptor ROR1
UniProt Gene Name	ROR1
UniProt Synonym Gene Names	NTRKR1
UniProt Entry Name	ROR1_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.