

Human keyhole limpet hemocyanin (KLH) antibody (IgG) ELISA Kit

Human keyhole limpet hemocyanin (KLH) antibody, IgG ELISA Kit

CATALOG #
AAA15814

REACTIVITY
Human

PRODUCT SYNONYMS

keyhole limpet hemocyanin (KLH) antibody (IgG); Human keyhole limpet hemocyanin (KLH) antibody; IgG ELISA Kit; Human keyhole limpet hemocyanin (KLH) antibody (IgG) ELISA Kit; keyhole limpet hemocyanin (KLH) antibody (IgG) elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of human HSVI antibody (IgG). No significant cross-reactivity or interference between human HSVI antibody (IgG) and analogues was observed.
Samples	Serum
Intra-assay Precision	Intra-assay Precision (Precision within an assay): CV%<15%. Three samples of known concentration were tested twenty times on one plate to assess
Inter-assay Precision	Inter-assay Precision (Precision between assays): CV%<15%. Three samples of known concentration were tested in twenty assays to assess.
Preparation and Storage	Unopened test kits should be stored at 2 to 8 degree C upon receipt. Please refer to pdf manual for further storage instructions.

IMAGES

TYPICAL DATA		
Test parameter	specification	test result
Positive control	>0.8	1.277
Negative control	<0.3	0.141
Positive rate	10, Positive	100%
Negative rate	10, Negative	100%

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for keyhole limpet hemocyanin (KLH) antibody (IgG) elisa kit

Principle of the Assay: This assay employs the qualitative enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with antigen. Samples are pipetted into the wells with anti-human IgG conjugated Horseradish Peroxidase (HRP). Any antibodies specific for the antigen present will bind to the pre-coated antigen. Following a wash to remove any unbound reagent, a substrate solution is added to the wells and color develops in proportion to the amount of human HSVI antibody (IgG) bound in the initial step. The color development is stopped and the intensity of the color is measured.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.