

Human Inhibin Beta E (INHbE) CLIA Kit | INHbE clia kit

Human Inhibin Beta E (INHbE) CLIA Kit

CATALOG #
AAA161500

REACTIVITY
Human

PRODUCT SYNONYMS

Inhibin Beta E (INHbE); Human Inhibin Beta E (INHbE) CLIA Kit; INHbE CLIA kit (Human); INH-bE; Activin beta-E chain; INHbE clia kit

PRODUCT SPECIFICATIONS


Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Inhibin Beta E (INHbE). No significant cross-reactivity or interference between Inhibin Beta E (INHbE) and analogues was observed. Reactive with Human INHbE/Inhibin Beta E
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma, Tissue homogenates, Cell lysates, Cell culture supernates and other biological fluids
Detection Method	Chemiluminescence
Featured Series	SE kit
Detection Range	1.37-1,000pg/mL
Sensitivity	0.54pg/mL
Sample Size	100uL
Assay Time	2h, 40min
Function	Detects protein (regular version)
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Inhibin Beta E (INHbE) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Inhibin Beta E (INHbE) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10%, Inter-Assay: CV<12%
Assay Procedure Summary	1. Prepare all reagents, samples and standards 2. Add 100ul standard or sample to each well. Incubate 2 hours at 37 degree C 3. Aspirate and add 100ul prepared Detection Reagent A. Incubate 1 hour at 37 degree C 4. Aspirate and wash 3 times 5. Add 100ul prepared Detection Reagent B. Incubate 30 minutes at 37 degree C 6. Aspirate and wash 5 times 7. Add 100ul Substrate Solution. Incubate 10 minutes at 37 degree C 8. Read RLU value immediately.
Test Principle	The microplate provided in this kit has been pre-coated with an antibody specific to Inhibin Beta E (INHbE). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Inhibin Beta E (INHbE). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Inhibin Beta E (INHbE) level in the sample or standard.
Research Area	Endocrinology

Preparation and Storage

Store at 4 degree C for 1 month/-20 degree C for long-term (One year within shelf life)

The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

Typical Standard Curve for INHbE CLIA (Sandwich)

NCBI AND UNIPROT INFORMATION

NCBI GI #	13899338
NCBI GeneID	83729
NCBI Accession #	NP_113667.1
NCBI GenBank Nucleotide #	NM_031479.3
UniProt Accession #	P58166
Molecular Weight	38,561 Da
NCBI Official Full Name	inhibin beta E chain
NCBI Official Synonym Full Names	inhibin, beta E
NCBI Official Symbol	INHBE
NCBI Protein Information	inhibin beta E chain; activin beta E; activin beta-E chain
UniProt Protein Name	Inhibin beta E chain
UniProt Gene Name	INHBE

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/human-clia-kits/inhibin-beta-e-inhbe/161500>