

Pig Complement Component 7 (C7) ELISA Kit | C7 elisa kit

Pig Complement Component 7 (C7) ELISA Kit

CATALOG #
AAA161542

REACTIVITY
Pig

PRODUCT SYNONYMS


Complement Component 7 (C7); Pig Complement Component 7 (C7) ELISA Kit; C7 ELISA kit (Pig); C7 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Pig
Specificity	This assay has high sensitivity and excellent specificity for detection of Complement Component 7 (C7). No significant cross-reactivity or interference between Complement Component 7 (C7) and analogues was observed. Reactive with Pig C7/Complement Component 7
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma and other biological fluids
Detection Method	Colormetric
Featured Series	SE kit
Detection Range	31.2-2,000ng/mL
Sensitivity	15.4ng/mL
Sample Size	100uL
Assay Time	3h
Function	Detects protein (regular version)
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Complement Component 7 (C7) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Complement Component 7 (C7) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10%, Inter-Assay: CV<12%
Assay Procedure Summary	1. Prepare all reagents, samples and standards 2. Add 100ul standard or sample to each well. Incubate 2 hours at 37 degree C 3. Aspirate and add 100ul prepared Detection Reagent A. Incubate 1 hour at 37 degree C 4. Aspirate and wash 3 times 5. Add 100ul prepared Detection Reagent B. Incubate 30 minutes at 37 degree C 6. Aspirate and wash 5 times 7. Add 90ul Substrate Solution. Incubate 10-20 minutes at 37 degree C 8. Add 50ul Stop Solution. Read at 450nm immediately.
Test Principle	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Complement Component 7 (C7). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Complement Component 7 (C7). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Complement Component 7 (C7), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm +/- 10nm. The concentration of Complement Component 7 (C7) in the samples is then determined by comparing the O.D. of the samples to the standard curve.

Research Area	Infection immunity, Immune molecule
Preparation and Storage	Store at 4 degree C for 1 month/-20 degree C for long-term (One year within shelf life) The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)
Typical Standard Curve for C7 ELISA (Sandwich)

NCBI AND UNIPROT INFORMATION

NCBI GI #	47523630
NCBI GeneID	397526
NCBI Accession #	NP_999447.1
NCBI GenBank Nucleotide #	NM_214282.1
UniProt Accession #	Q9TUQ3
Molecular Weight	93,062 Da
NCBI Official Full Name	complement component C7
NCBI Official Symbol	C7
NCBI Protein Information	complement component C7
UniProt Protein Name	Complement component C7
UniProt Gene Name	C7

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
 Online datasheet: <https://stage.aaabiotech.com/pig-elisa-kits/complement-component-7-c7/161542>