# Human FcgammaR3A ELISA Kit | FcgammaR3A elisa kit

Human FcgammaR3A (Fc Fragment of IgG Low Affinity IIIa Receptor) ELISA Kit

| CATALOG # | REACTIVITY |
|---|---|
| **AAA164011** | **Human** |

## PRODUCT SYNONYMS

FcgammaR3A; Human FcgammaR3A (Fc Fragment of IgG Low Affinity IIIa Receptor) ELISA Kit; FcgammaR3A elisa kit

## PRODUCT SPECIFICATIONS

| | |
|---|---|
| **Reactivity** | Human |
| **Specificity** | This assay has high sensitivity and excellent specificity for detection of Human FcgammaR3A. No significant cross-reactivity or interference between Human FcgammaR3A and analogues was observed. Note: Limited by current skills and knowledge, it is impossible for us to complete the cross- reactivity detection between Human FcgammaR3A and all the analogues, therefore, cross reaction may still exist. |
| **Samples** | Serum, Plasma, Biological Fluids |
| **Assay Type** | Sandwich |
| **Detection Range** | 7.813-500ng/mL |
| **Sensitivity** | Min: 4.688ng/mL; Max: 500ng/mL |
| **Preparation and Storage** | Store at 4 degree C. |

## IMAGES



**Standard Curve (Sample)**

## ADDITIONAL INFORMATION

| | |
|---|---|
| **Related Product Information for FcgammaR3A elisa kit** | **Description:** The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of FcgammaR3A in human serum, plasma and other biological fluids<br>**Principle of the Assay:** This ELISA kit uses Sandwich-ELISA as the method. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human FcgammaR3A. Standards or samples are added to the appropriate micro ELISA plate wells and bound by the specific antibody. Then a biotinylated detection antibody specific for Human FcgammaR3A and Avidin-Horseradish Peroxidase (HRP) conjugate is added to each micro plate well successively and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human FcgammaR3A, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of a sulphuric acid solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm +/- 2 nm. The OD value is proportional to the concentration of Human FcgammaR3A. You can calculate the concentration of Human FcgammaR3A in the samples by comparing the OD of the samples to the standard curve. |

**For Research Use Only.** Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

**AAA Biotech** | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: https://stage.aaabiotech.com/human-elisa-kits/fcgammar3a/164011