

Human Anti Myelin Oligodendrocyte Glycoprotein Antibody ELISA Kit | MOG elisa kit

Human Anti Myelin Oligodendrocyte Glycoprotein Antibody ELISA Kit

CATALOG #
AAA16409

REACTIVITY
Human


PRODUCT SYNONYMS

Anti Myelin Oligodendrocyte Glycoprotein Antibody; Human Anti Myelin Oligodendrocyte Glycoprotein Antibody ELISA Kit; MOG elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Anti-MOG. No significant cross-reactivity or interference between Anti-MOG and analogues was observed. NOTE: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between Anti-MOG and all the analogues, therefore, cross reaction may still exist in some cases.
Assay Type	Quantitative Competitive or Sandwich
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Sensitivity	1.0 ng/mL
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for MOG elisa kit	Principle of the Assay: Anti-MOG ELISA kit applies the quantitative sandwich enzyme immunoassay technique. The microtiter plate has been pre-coated with Myelin Oligodendrocyte Glycoprotein antigen. Standards or samples are then added to the microtiter plate wells and Anti-MOG if present, will bind to the Myelin Oligodendrocyte Glycoprotein antigen pre-coated wells. In order to quantitatively determine the amount of Anti-MOG present in the sample, a standardized preparation of horseradish peroxidase (HRP)-conjugated Anti-MOG is added to each well to "sandwich" the Anti-MOG immobilized on the plate. The microtiter plate undergoes incubation, and then the wells are thoroughly washed to remove all unbound components. Next, substrate solutions are added to each well. The enzyme (HRP) and substrate are allowed to react over a short incubation period. Only those wells that contain Anti-MOG and enzyme-conjugated antibody will exhibit a change in color. The enzyme-substrate reaction is terminated by addition of a sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450 nm. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The Anti-MOG concentration in each sample is interpolated from this standard curve. Intended Uses: This Anti-MOG ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Human Anti-MOG. This ELISA kit for research use only, not for therapeutic or diagnostic applications!
Product Categories/Family for MOG elisa kit	Cell Biology

NCBI AND UNIPROT INFORMATION

NCBI GI #	168985765
-----------	-----------

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/mog-human-elisa-kits/anti-myelin-oligodendrocyte-glycoprotein-antibody/16409>