

Rat OLAB ELISA Kit | OLAB elisa kit

Rat OLAB (Oxidized Lowdensity Lipoprotein Antibody) ELISA Kit

CATALOG #
AAA164153

REACTIVITY
Rat


PRODUCT SYNONYMS

OLAb; Rat OLAB (Oxidized Lowdensity Lipoprotein Antibody) ELISA Kit; OLAB elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This kit recognizes natural and recombinant Rat OLAB. No significant cross-reactivity or interference between Rat OLAB and analogues was observed.
Samples	Serum, Plasma, Biological Fluids
Assay Type	Sandwich
Detection Range	3.125-200ng/mL
Sensitivity	Min: 1.875ng/mL; Max: 200ng/mL
Preparation and Storage	Store at 4 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for OLAB elisa kit

Intended Uses: This ELISA kit applies to the in vitro quantitative determination of Rat OLAB concentrations in serum, plasma and other biological fluids.

Principle of the Assay: This ELISA kit uses Sandwich-ELISA as the method. The micro ELISA plate provided in this kit has been pre-coated with an antigen specific to OLAB. Standards or samples are added to the appropriate micro ELISA plate wells and combined with the specific antigen. Then a biotinylated detection antigen specific for OLAB and Avidin-Horseradish Peroxidase (HRP) conjugate is added to each micro plate well successively and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain OLAB, biotinylated detection antigen and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of a sulphuric acid solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm $\pm$ 2 nm. The OD value is proportional to the concentration of OLAB. You can calculate the concentration of OLAB in the samples by comparing the OD of the samples to the standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/rat-elisa-kits/olab/164153>