

Human OLAB ELISA Kit | OLAB elisa kit

Human OLAB (Anti-Oxidized Low Density Lipoprotein Antibody) ELISA Kit

CATALOG #
AAA164455

REACTIVITY
Human


PRODUCT SYNONYMS

OLAb; Human OLAB (Anti-Oxidized Low Density Lipoprotein Antibody) ELISA Kit; OLAB elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Human OLAB. No significant cross-reactivity or interference between Human OLAB and analogues was observed. Note: Limited by current skills and knowledge, it is impossible for us to complete the cross- reactivity detection between Human OLAB and all the analogues, therefore, cross reaction may still exist.
Samples	Serum, Plasma, Biological Fluids
Assay Type	Sandwich
Detection Range	0.938-60mIU/mL
Sensitivity	Min: 0.563mIU/mL; Max: 60mIU/mL
Preparation and Storage	Store at 4 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for OLAB elisa kit	Description: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of OLAB in human serum, plasma and other biological fluids Principle of the Assay: This ELISA kit uses Sandwich-ELISA as the method. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human OLAB. Standards or samples are added to the appropriate micro ELISA plate wells and bound by the specific antibody. Then a biotinylated detection antibody specific for Human OLAB and Avidin-Horseradish Peroxidase (HRP) conjugate is added to each micro plate well successively and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human OLAB, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of a sulphuric acid solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm +/- 2 nm. The OD value is proportional to the concentration of Human OLAB. You can calculate the concentration of Human OLAB in the samples by comparing the OD of the samples to the standard curve.
--	---

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/olab/164455>