

Rat beta-EPR ELISA Kit | beta-EPR elisa kit

Rat beta-EPR (Beta-Endorphin Receptor) ELISA Kit

CATALOG #
AAA165116

REACTIVITY
Rat


PRODUCT SYNONYMS

beta-EPR; Rat beta-EPR (Beta-Endorphin Receptor) ELISA Kit; beta-EPR elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This kit recognizes natural and recombinant Rat beta-EPR. No significant cross-reactivity or interference between Rat beta-EPR and analogues was observed.
Samples	Serum, Plasma, Biological Fluids
Assay Type	Sandwich
Detection Range	31.25-2000pg/mL
Sensitivity	Min: 18.75pg/mL; Max: 2000pg/mL
Preparation and Storage	Store at 4 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for beta-EPR elisa kit

Intended Uses: This ELISA kit applies to the invitro quantitative determination of Rat beta-EPR concentrations in serum, plasma and other biological fluids.

Principle of the Assay: This ELISA kit uses Sandwich-ELISA as the method. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to beta-EPR. Standards or samples are added to the appropriate micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for beta-EPR and Avidin-Horseradish Peroxidase (HRP) conjugate is added to each micro plate well successively and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain beta-EPR, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of a sulphuric acid solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm +/- 2 nm. The OD value is proportional to the concentration of beta-EPR. You can calculate the concentration of beta-EPR in the samples by comparing the OD of the samples to the standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/rat-elisa-kits/beta-epr/165116>