

Rabbit Vitamin D ELISA Kit | VD elisa kit

Rabbit Vitamin D ELISA Kit

CATALOG #
AAA16850

REACTIVITY
Rabbit


PRODUCT SYNONYMS

Vitamin D; Rabbit Vitamin D ELISA Kit; VD elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rabbit
Specificity	This assay has high sensitivity and excellent specificity for detection of VD. No significant cross-reactivity or interference between VD and analogues was observed. NOTE: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between VD and all the analogues, therefore, cross reaction may still exist in some cases.
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Quantitative Competitive
Sensitivity	1.0 ng/mL
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for VD elisa kit	Intended Uses: This VD ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Rabbit VD. This ELISA kit for research use only! Principle of the Assay: VD ELISA kit applies the competitive enzyme immunoassay technique utilizing a polyclonal anti-VD antibody and an VD-HRP conjugate. The assay sample and buffer are incubated together with VD-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the VD concentration since VD from samples and VD-HRP conjugate compete for the anti-VD antibody binding site. Since the number of sites is limited, as more sites are occupied by VD from the sample, fewer sites are left to bind VD-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The VD concentration in each sample is interpolated from this standard curve.
--	---

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/vd-rabbit-elisa-kits/vitamin-d/16850>