

Human beta-EPR (Beta-Endorphin Receptor) CLIA Kit | b-EPR clia kit

Human beta-EPR (Beta-Endorphin Receptor) CLIA Kit

CATALOG #
AAA170861

REACTIVITY
Human


PRODUCT SYNONYMS

beta-EPR (Beta-Endorphin Receptor); Human beta-EPR (Beta-Endorphin Receptor) CLIA Kit; b-EPR; Human b-EPR (Beta-Endorphin Receptor) CLIA Kit; Human b-EPR; b-EPR clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This kit recognizes natural and recombinant beta-EPR. No significant cross-reactivity or interference between beta-EPR and analogues was observed.
Assay Type	Sandwich
Detection Range	12.5-800pg/mL
Sensitivity	Min: 7.5pg/mL; Max: 800pg/mL

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for b-EPR clia kit

Intended Uses: This CLIA kit applies to the in vitro quantitative determination of beta-EPR concentrations in serum, plasma and other biological fluids.

Principle of the Assay: This kit uses Sandwich-CLIA as the method. The micro CLIA plate provided in this kit has been pre-coated with an antibody specific to beta-EPR. Standards or samples are added to the appropriate micro CLIA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for beta-EPR and Avidin-Horseradish Peroxidase (HRP) conjugate is added to each micro plate well successively and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain beta-EPR, biotinylated detection antibody and Avidin-HRP conjugate will appear fluorescence. The Relative light unit (RLU) value is measured spectrophotometrically by the Chemiluminescence immunoassay analyzer. The RLU value is positively associated with the concentration of beta-EPR. You can calculate the concentration of beta-EPR in the samples by comparing the RLU value of the samples to the standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

