

Canine Resolvin E1 ELISA Kit | RvE1 elisa kit

Canine Resolvin E1 ELISA Kit

CATALOG #
AAA17089

REACTIVITY
Canine


PRODUCT SYNONYMS

Resolvin E1; Canine Resolvin E1 ELISA Kit; RvE1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Canine
Specificity	This assay has high sensitivity and excellent specificity for detection of RvE1. No significant cross-reactivity or interference between RvE1 and analogues was observed. NOTE: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between RvE1 and all the analogues, therefore, cross reaction may still exist in some cases.
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Quantitative Competitive
Sensitivity	0.1 ?g/mL
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for RvE1 elisa kit	Intended Uses: This RvE1 ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Canine RvE1. This ELISA kit for research use only, not for therapeutic or test applications! Principle of the Assay: RvE1 ELISA kit applies the competitive enzyme immunoassay technique utilizing a polyclonal anti-RvE1 antibody and an RvE1-HRP conjugate. The assay sample and buffer are incubated together with RvE1-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the RvE1 concentration since RvE1 from samples and RvE1-HRP conjugate compete for the anti-RvE1 antibody binding site. Since the number of sites is limited, as more sites are occupied by RvE1 from the sample, fewer sites are left to bind RvE1-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The RvE1 concentration in each sample is interpolated from this standard curve.
--	---

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/rve1-canine-elisa-kits/resolvin-e1/17089>